

TEST REPORT: 7191247377-CHM20-01-RC
12 NOV 2020



PSB Singapore

RESULTS

Sample Description: Enro – Tech 3D Mask (before wash)

Test Method:
ASTM F2101

Test sample/ controls	Stage 1, CFU	Stage 2, CFU	Stage 3, CFU	Stage 4, CFU	Stage 5, CFU	Stage 6, CFU	Sum of Total plate count for the 6 sieves, CFU	Average Count for Controls, CFU	BFE (%)
-ve Control	0	0	0	0	0	0	Less than 1		
+ve Control 1	108	343	543	530	389	14	1927	1928	
+ve Control 2	98	324	567	568	364	7	1928		
Sample 1	0	0	0	0	0	0	Less than 1		More than 99.9
Sample 2	0	0	0	0	0	0	Less than 1		More than 99.9
Sample 3	0	0	0	0	0	0	Less than 1		More than 99.9
Sample 4	0	0	0	0	0	0	Less than 1		More than 99.9
Sample 5	0	0	0	0	0	0	Less than 1		More than 99.9